

RESEARCH ARTICLE


Epitope mapping of recombinant *Leishmania donovani* virulence factor A2 (recLdVFA2) and canine leishmaniasis diagnosis using a derived synthetic bi-epitope

Thais Melo Mendes¹*, Eric Henrique Roma¹*, Fernanda Costal-Oliveira¹, Lucas de Carvalho Dhom-Lemos², Cristina Monerat Toledo-Machado², Oscar Bruna-Romero³, Daniella Castanheiras Bartholomeu², Ricardo Toshio Fujiwara², Carlos Chávez-Olórtegui¹*

1 Departamento de Bioquímica e Imunologia, Instituto de Ciências Biológicas, UFMG, Belo Horizonte—Minas Gerais, Brazil, **2** Departamento de Parasitologia, Instituto de Ciências Biológicas, UFMG, Belo Horizonte—Minas Gerais, Brazil, **3** Departamento de Microbiologia, Imunologia e Parasitologia, CCB, UFSC, Florianópolis—Santa Catarina, Brazil

* These authors contributed equally to this work.

* olortegi@icb.ufmg.br



OPEN ACCESS

Citation: Mendes TM, Roma EH, Costal-Oliveira F, Dhom-Lemos LdC, Toledo-Machado CM, Bruna-Romero O, et al. (2017) Epitope mapping of recombinant *Leishmania donovani* virulence factor A2 (recLdVFA2) and canine leishmaniasis diagnosis using a derived synthetic bi-epitope. PLoS Negl Trop Dis 11(5): e0005562. <https://doi.org/10.1371/journal.pntd.0005562>

Editor: Hechmi Louzir, Institut Pasteur de Tunis, TUNISIA

Received: February 16, 2016

Accepted: April 11, 2017

Published: May 30, 2017

Copyright: © 2017 Mendes et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: All relevant data are within the paper and its Supporting Information files.

Funding: CCO and TMM were supported in part by – Coordenação de Aperfeiçoamento de Pessoal de Nível Superior - CAPES [27639/2009-57 and 02448/09-2], [HTTP://www.capes.gov.br](http://www.capes.gov.br). This work also has the support of Fundação de Amparo à Pesquisa do Estado de Minas Gerais – FAPEMIG

Abstract

Background

Leishmaniasis is one of the most important zoonotic diseases spread in Latin America. Since many species are involved in dog infection with different clinical manifestations, the development of specific diagnostic tests is mandatory for more accurate disease control and vaccine strategies.

Methodology/Principal findings

Seventy-five 15-mer peptides covering the sequence of recombinant *Leishmania donovani* virulence factor A2 (recLdVFA2) protein were prepared by Spot synthesis. Membrane-bound peptides immunoreactivity with sera from dogs immunized with recLdVFA2 and with a specific anti-recLdVFA2 monoclonal antibody allowed mapping of continuous B-cell epitopes. Five epitopes corresponding to the N-terminal region of recLdVFA2 (MKIRSVRPLVLLVC, RSV RPLVLLVCVAA, RPLVLLVCVAVLA, VLLVCVAVLALSA and LVCVAVLALSASAE, region 1–28) and one located within the repetitive units (PLSVGPQAVGLSVG, regions 67–81 and 122–135) were identified. A 34-mer recLdVFA2-derived bi-epitope containing the sequence MKIRSVRPLVLLVC linked to PLSVGPQAVGLSVG by a Gly-Gly spacer was chemically synthesized in its soluble form. The synthetic bi-epitope was used as antigen to coat ELISA plates and assayed with dog sera for in vitro diagnosis of canine visceral leishmaniasis (CVL). The assay proved to be highly sensitive (98%) and specific (99%).

Conclusions/Significance

Our work suggests that synthetic peptide-based ELISA strategy may be useful for the development of a sensitive and highly specific serodiagnosis for CVL or other parasitic diseases.

[CBB-APQ 00831-15], [HTTP://www.fapemig.br](http://www.fapemig.br) and of Conselho Nacional de Desenvolvimento Científico e Tecnológico - CNPq [CNPq number 407266], [HTTP://www.cnpq.br](http://www.cnpq.br). The funders had no role in study design, data collection and analysis decision to publish, or preparation on the manuscript.

Competing interests: The authors have declared that no competing interest exist.

Author summary

Leishmaniasis is a neglected tropical disease being among the six endemic prioritized diseases in the world. Visceral leishmaniasis (VL) is caused by *Leishmania infantum* and represents a serious public health problem in Brazil. Dogs are the main source of infection in the urban area and, in Brazil, the main strategies of the Visceral Leishmaniasis Control Program are directed to control the canine reservoir (serological survey and euthanasia of dogs which present reactive serum). In general, diagnosis of canine visceral leishmaniasis (CVL) has been presented as a problem for Brazilian public health services. The issue should be attributed mainly to the following factors: 1- range of similar clinical signs observed in other infectious diseases that affect dogs; 2- large percentage of asymptomatic or oligosymptomatic dogs; 3- nonspecific histopathological changes; 4- nonexistence of a diagnostic test 100% specific and sensitive. In this work, we developed a synthetic bi-epitope peptide as an antigen for immunodiagnostic ELISA to detect CVL. The bi-epitope used for ELISA assay accurately distinguish (98% sensitivity and 99% specificity) CVL dogs sera from non-infected dogs sera.

Introduction

Visceral leishmaniasis (VL) is an infection caused by various species of *Leishmania*, an intracellular protozoan parasite. Currently, VL is among the six endemic prioritized diseases in the world [1]. In humans, infection with *Leishmania* can cause a broad spectrum of symptoms ranging from a clinically silent infection to a fatal visceral disease [2]. In an urban environment, dogs are the main reservoir of the disease, but many stay asymptomatic, showing no clinical signs [3, 4, 5].

A2 is a stress response protein from *L. donovani* and it is expressed in amastigote and in promastigote cultures. It corresponds to the specific virulence factor (*LdVFA2*) and has been shown to be required for *L. donovani* amastigote survival in visceral organs of mice [6, 7, 8]. A2 proteins are composed mostly of a variable number of 10-amino-acid repeats and their molecular weight varies from 45 to 100 kDa [9]. *LdVFA2* antigens, administered as recombinant protein (*recLdVFA2*) or DNA, are protective against *L. donovani*, *L. amazonensis* and *L. chagasi* infections in mice [10, 11, 12], dogs [13] and macaques [14]. Anti-*LdVFA2* antibodies have been detected in sera samples from human patients with active visceral leishmaniasis, confirming that *LdVFA2* proteins are expressed during infection [10, 15]. These findings suggest that studies of *LdVFA2* proteins antigenic properties might have great potential for the development of vaccines, therapeutics and diagnostics for leishmaniasis.

In this work, we report the mapping of B-cell continuous epitopes of *recLdVFA2*, production by chemical synthesis of a *recLdVFA2*-derived synthetic epitope and its use as antigen for canine visceral leishmaniasis (CVL) diagnosis. Epitope mapping was achieved by peptide-scanning of the *recLdVFA2* sequence using the Spot-synthesis technique [16]. This method is an easy and very flexible technique for simultaneous parallel peptides chemical synthesis on membrane supports. Furthermore, it allows a rapid and low-cost access to a large number of peptides for systematic epitope analysis [17]. Sixty-five overlapping peptides (15-mer frame-shifted by 3 residues) covering the complete amino acid sequence of *recLdVFA2* were synthesized on cellulose membranes. Five continuous epitopes corresponding to the non-repetitive N-ter region of *recLdVFA2* (MKIRSVRPLVLLVC, RSVRPLVLLVCVAA, RPLVLLVC VAAVLA, VLLVCVAAVLALSA and LVCVAAVLALSASAE, region 1–28) were mapped

using anti- *recLdVFA2* dog sera and one epitope was located within the repetitive units (PLSVGPQAVGLSVG, region 67–81 and 122–135) using an anti-*recLdVFA2* mAb. An epitope from N-ter (MKIRSVRPLVLLVC) and another from the C-ter part (PLSVGPQAVGLSVG) were selected and chemically assembled in tandem, to yield a soluble bi-epitope peptide. The bi-epitope used as coating antigen in ELISA accurately distinguish (high sensitivity and specificity) sera of CVL dogs from sera of non-infected dogs.

Methods

In order to discover a new antigen for use in CVL diagnostics tests, dog's polyclonal and monoclonal anti-sera *LdVFA2* were used to map epitopes. An initial screening was made to select polyclonal antisera with high affinity for *LdVFA2* protein and low affinity for the crude extract of *Leishmania infantum*. The epitopes mapped to the selected polyclonal sera and with the specific monoclonal sera were analyzed and two were selected and synthesized as a bi-peptide. This bi-epitope peptide was called *recLdVFA2* derived and it was tested as antigen for CVL diagnosis. The choice of *L. donovani* protein for epitope search for use in *L. infantum* diagnostic test is due to extensive research on this protein (*recVFA2*), amino acid sequence availability and biological findings indicating that there is interspecific cross-reactivity.

Ethics statement

Approval to use the sera samples was obtained from the Committee on Ethics of Animal Experimentation (CETEA, national guidelines Lei 11.794, de 8 de outubro de 2008) from this UFMG (CETEA–protocol 44/2012).

Production *leishmania (leishmania) infantum* antigen (*LiA*) and *recLdVFA2*

L. infantum (MHOM/BR/1975/BH46) was grown at 24°C in Schneider's medium (Sigma, St. Louis, MO, USA) supplemented with 20% heat-inactivated fetal bovine serum (FBS; Sigma), 200 U/mL penicillin and 100 µg/mL streptomycin, at pH 7.2. Total soluble antigens of *L. infantum* (*LiA*) was prepared from stationary phase promastigotes, submitted to 7 cycles of freezing (liquid nitrogen) and thawing (42°C), followed by ultrasonication (Ultrasonic processor, GEX600), with cycles of 10 sec for 2 min at 35 MHz. Extracts were then submitted to centrifugation at 8000 × *g* for 20 min at 4°C. The supernatant was collected and stored at –70°C. *recLdVFA2*, a recombinant form of A2, was expressed and purified as previously described [15].

Dog sera and monoclonal antibody against *recLdVFA2*

For epitope mapping, sera from seventy-three dogs immunized with *recLdVFA2* were obtained from the laboratory HERTAPE-KALIER Health Animal S.A. For diagnostic test, serum of dogs from a CVL non-endemic area and giving negative results for *L. infantum* in immunofluorescence antibody test (IFAT) and confirmed by parasitological test and microscopic analysis of bone marrow aspirates were considered to be non-infected and used as the control group (NI, n = 101). *Leishmania*-infected dog sera (I, n = 101) were obtained from an endemic area for CVL in the Minas Gerais State of Brazil. Infection status was determined by parasitological test, the positivity was confirmed by microscopic analysis of bone marrow aspirates. The positive and negative status was further confirmed by real time PCR. Samples from dogs experimentally infected with *Trypanosoma cruzi* (TC, n = 10), but parasitologically negative for *Leishmania*, were included in this study to evaluate possible cross-reactivity. All sera

used for diagnostic test were obtained from sera bank already existing of Laboratory of Immunology and Genomics of Parasites, Federal University of Minas Gerais (UFMG/BR). The anti-*L. donovani* virulence factor A2 murine monoclonal antibody (mAb-anti recLdVFA2) was kindly provided by Dr. Greg Matlashewski, McGill University, Quebec, Canada.

Identification new antigen for CVL diagnostic

Selection of polyclonal sera by ELISA. Falcon flexible microtitration plates (Becton Dickinson France S.A.) were coated overnight at 5°C with 100 µl of a 10 µg/ml solution of *LiA* and 2 µg/ml solution of recLdVFA in 0.02M sodium bicarbonate buffer, pH 9.6. Plates were washed with PBS 1x, 0.1% Tween 20 (v/v) and then blocked with PBS 1x, 0.1% Tween 20 and 2% caseine (w/v) for 1 h at 37°C. Anti-recLdVFA dog sera (diluted 1:100) were added to respective wells. Plates were incubated for 2 h at 37°C. Binding was detected using an anti-Dog IgG peroxidase conjugated (Sigma) diluted 1:5000 in blocking buffer. After 1 h at 37°C and washing, the peroxidase substrate was added. The reaction was stopped with H₂SO₄ 1N after 30 minutes and the resulting colour was measured at 492 nm with an automated microtiter plate reader (Model 450, Bio-Rad). All measurements were made in triplicate.

B-cell epitope mapping of recLdVFA

Spot peptides synthesis on cellulose membrane. Seventy-five 15-mer overlapping peptide frameshifted by three amino acids derived from the sequence of A2 protein (GenBank accession number: AAB30592.1) were synthesized on a cellulose membrane (Intavis, Koln, Germany) according to [18] and an ASP222 robot (Intavis) was used for the coupling steps. N-terminal acetylation of the peptides was also performed to increase their stability. After peptide sequences had been assembled, the side-chain protecting groups were removed by trifluoroacetic acid treatment as described before [19].

Immunoassay with cellulose membrane-bound peptides. After an overnight saturation step with blocking buffer (Genosys, France), the set of membrane bound peptides was probed by incubation with serum from four dogs immunized with recA2p (diluted 1:100) for 2 h. Polyclonal antibodies binding was detected by using an alkaline phosphatase-conjugated anti-dog antibody (Sigma) diluted 1:2000 for 1 h, and by the addition of a phosphatase substrate [5-bromo-4-chloro-3-indolyl phosphate (BCIP) and 3-(4,5-dimethylthiazol-2-yl)-2,5-diphenyltetrazolium bromide (MTT) (Sigma), which generates a blue precipitate over spots harboring a peptide recognized by antibodies. A similar procedure was adopted for assays with mAb, used at 2.0 µg/ml. In this case, peptide reactivity was revealed by using an alkaline phosphatase-conjugated anti-mouse antibody (Sigma) diluted 1:2000. To allow its reuse, the membrane was sequentially treated with dimethylformamide, 1% SDS, 0.1% 2-mercaptoethanol in 8 M urea, ethanol/water/acetic acid (50:40:10 vol/vol/vol) and, finally, methanol to remove the precipitated dye and molecules bound to the peptides. The intensity of spots colors were arbitrary quantified using ImageJ software (NIH, Bethesda, MD, USA). The highest reactive spot for each type of epitope mapping (spot 3, for mapping using polyclonal immunized dog sera and spot 48 for mapping using mAb) was considered as presenting 100% of reactivity. The reactivity of all other spots was expressed as percentage of values according to the highest reactive spots.

Bi-epitope synthesis of recLdVFA2 derived. Based on the results of peptide array immunoassay, it was prepared a 34-mer peptide containing two sequences of fifteen amino acids residues in tandem: sequence PLSVGPQAVGLSVG from repetitive part of A2 protein was linked by a Gly-Gly dipeptide to the 15-mer sequence MKIRSVRPLVLLVC from the non-repetitive part, in order to bring together regions that are apart from each other in the linear sequence of

the target peptide. With the purpose of facilitate subsequent affinity based purification of the anti-peptide antibodies, two amino acid residues were added, Lys to the N-ter and Cys to C-ter extremities, respectively. The 34-mer peptide was synthesized manually by F-moc chemistry on a Wang resin (Novabiochem) as described [20] adapted by [21]. After synthesis the protecting groups of side chains were deprotected and the peptide was released from the resin by using TFA in the presence of appropriate scavengers. Then, the peptide was lyophilized and its mass confirmed by mass spectrometry. The synthetic bi-epitope *recLdVFA2* derived (20 mg) was diluted in 1 ml of PBS and used as antigenic preparation to coat ELISA plates. A non-related peptide CRCKPDQGRLRCGYK (nrpep) was also synthesized to be used as negative control.

Diagnosis of canine leishmaniasis using synthetic bi-epitope

One hundred and one sera from infected dogs were used (I), as well as 101 sera from dogs without a history of *Leishmania* infections (NI) and 10 sera from dogs experimentally infected by *T. cruzi* (TC), to verify if the bi-epitope *recLdVFA2* derived is a good candidate to be antigen in CVL diagnostic test. Maxisorb flexible microtitration plates were coated overnight at 5°C with 100 µL of synthetic peptide solution (10 µg/mL) in 0.02 M sodium bicarbonate buffer, pH 9.6. Assays were performed as previously described [22]. Sera were diluted 1:100 and absorbance values were determined at 492 nm with a Titertek Multiscan spectrophotometer. All measurements were made in triplicate. Standard EIE-LCV kit for the leishmaniasis diagnosis was used for comparison. This test is the most used in the clinical and serologic testing ELISA for LCV is a good test for use in the field epidemiological serum screening due to its convenience and low cost.

Statistical analysis

All data comparisons were tested for significance by using unpaired Student's *t* test or Kruskal-Wallis test. Differences were considered statistically significant when *P* values were < 0.05. The lower limit of positivity (cut-off) for bi-epitope and EIE-LCV was established for optimal sensitivity and specificity using the Receiver Operator Curve (ROC curve). The cut-off was chosen based on the point that provides the maximum of the sum of the sensibility and specificity [23]. The performance of each test was evaluated according to the sensitivity (Se), specificity (Sp), positive predictive value (PPV), negative predictive value (NPV), area under the curve (AUC) and accuracy (ACC).

Statistical analyses were performed using GraphPad Prism (version 5.0) and R package for Windows (www.r-project.org) (version 3.1.0).

Results

Identification new antigen for CVL diagnostic

Selection of dog sera for B-cell epitope mapping. A study with 73 sera from dogs immunized with *recLdVFA2* protein was carried using ELISA to select sera for epitope mapping purposes. Data showed in [Table 1](#) indicate that anti-*recVFA2* antibodies were present in all tested sera samples (100%). ELISA values (Abs = 492 nm) varied between 0.195 and 1.4, with an average value of 0.721 ± 0.240 . Anti-*LiA* antibodies from dogs immunized with *recLdVFA2* were also measured. Sixty-nine of 73 sera samples, (94.5%) also exhibited cross-reactivity with *LiA*. ELISA values (Abs = 492 nm) varied between 0.055 and 1.10, with an average value of 0.410 ± 0.230 . Sera from dogs immunized with *recLdVFA2* which presented very low cross-reaction against *LiA* (Abs values less than 0.10) were used for the epitope mapping of *recLdVFA2*. Based in these

Table 1. ELISA reactivity of 73 sera of dogs immunized with recLdVFA2 against recLdVFA2 and LiA.

Dog serum number	Reactivity against		Dog serum number	Reactivity against	
	recLdVFA2	LiA		recLdVFA2	LiA
1	0.771	0.309	38	0.395	0.209
2	0.941	0.382	39	0.661	0.45
3	0.589	0.234	40	0.973	0.365
4	0.832	0.327	41	1.331	0.426
5	0.548	0.294	42	1.077	0.342
6	0.282	0.239	43	0.761	0.336
7	0.939	0.257	44	0.891	0.356
8	0.747*	0.069*	45	0.868*	0.065*
9	0.644	0.214	46	0.802	0.218
10	0.649	0.255	47	0.949	0.349
11	0.602	0.38	48	1.332*	0.095*
12	0.583	0.755	49	0.541	0.711
13	0.678	0.162	50	0.328	0.798
14	0.502	0.573	51	0.794	0.736
15	0.549	0.753	52	0.86	0.341
16	0.715	0.566	53	0.697	0.69
17	0.892	0.956	54	0.554	0.466
18	0.430	0.584	55	0.500	0.25
19	0.339	0.709	56	1.228*	0.095*
20	0.804	0.302	57	0.623	0.586
21	0.351	0.346	58	0.729	0.627
22	0.504	0.186	59	0.445	0.446
23	0.932	0.327	60	0.58	0.36
24	1.4	0.631	61	0.571	0.847
25	0.853	0.31	62	0.574	0.646
26	0.748	0.317	63	0.439	0.455
27	0.839	0.92	64	0.594	0.18
28	1.116	1.11	65	0.642	0.288
29	0.712	0.293	66	0.522	0.358
30	0.622	0.846	67	0.195	0.200
31	0.729	0.509	68	0.61	0.449
32	0.858	0.25	69	0.767*	0.056*
33	0.785	0.278	70	0.781	0.256
34	0.72	0.385	71	0.996	0.267
35	1.033	0.486	72	0.659	0.055
36	0.733	0.38	73	0.74	0.300
37	0.688	0.369			

ELISA plates were coated with of LiA and recLdVFA in 0.02M sodium bicarbonate buffer. Dog sera anti-recLdVFA were diluted 1:100 were added to respective wells. Peroxidase conjugated anti-IgG dog (Sigma) was 1:5000. The reactivity was measured at 492 nm with a microtiter plate reader (Model 450. Bio-Rad). All measurements were made in triplicate.

*Bold numbers indicate sera that gave a good reactivity response with the recLdVFA and a negative response with LiA.

<https://doi.org/10.1371/journal.pntd.0005562.t001>

criteria, sera from 4 dogs with highest ratio between the reactivity for recLdVFA2 and LiA were selected.

Identification of B-cell epitopes of recLdVFA2. To map the epitopes of recLdVFA2, it was used Spot-synthesis method [19, 24]. A set of 75 overlapping peptides (15 residues, offset by 3 residues) corresponding to the complete amino acid sequence of recLdVFA2 was synthesized in a cellulose membrane (Table 2). At the end of the synthesis, peptides remain covalently bound to the membrane and can be assayed for antibody reactivity. Four different sera from dogs immunized with recLdVFA2 and an anti-recLdVFA2 monoclonal antibody were used to probe the set of peptides for antigenic reactivity. Five peptides (MKIRSVRPLVVLLVC, RS VRPLVVLLVCVAA, RPLVVLLVCVAAVLA, VVLLVCVAAVLALSA and LVCVAAVLAL SASAE) corresponding to the non-repetitive N-terminal part of recLdVFA2 were strongly reactive (peptides 1–5) with all four anti-recVFA2 dog serum tested (Fig 1). Sera from non-immunized dogs were used as negative control showing no reactivity with membrane spots. When mAb anti-recLdVFA2 was used, spots corresponding to the non-repetitive N-ter region didn't react. However, all peptides of the repetitive C-ter region showed some degree of reactivity (Fig 2). Peptide number 48 (PLSVGPQAVGPLSVG) was the more strongly reactive (Table 2). Synthetic peptides of the repetitive region containing two replicates of the amino acid sequence PLSVG (peptides number 13, 23, 38, 48 and 52) showed high reactivity with the mAb and can be considered immunodominant epitopes of the repetitive region. However, some peptides containing only a single PLSVG motif also reacted quite strongly (peptides 27, 52, 75) with mAb.

Diagnosis of canine leishmaniasis using synthetic bi-epitope in an ELISA format

A 34-mer recVFA2-derived synthetic peptide containing the sequence MKIRSVRPLVVLLVC linked by a Gly-Gly to the sequence PLSVGPQAVGPLSVG was chemically synthesized. Two amino acids were added, a Lys to the N-ter and a Cys to the C-ter regions, respectively. This synthetic peptide (bi-epitope) was used as antigen to coat ELISA plates, for an immune diagnosis of CVL. ELISA parameters (eg. antigen concentration, incubation times, serum dilution) were previously defined.

A ready to use commercial EIE-LVC kit was included for performance comparison (Fig 3). In the conditions previously defined, the bi-epitope showed better diagnostic performance (AUC = 0.9987, 95% CI 0.9967 to 1.001; ACC = 0.9851) when compared to EIE-LVC kit (AUC = 0.9601, 95% CI 0.9351 to 0.9850; ACC = 0.9001) (Table 3). recLdVFA2 showed sensitivity (98.02%; 95% CI 93.03 to 99.76%) and specificity (99.01%; 95% CI 94.61 to 99.97%) values for detection CVL higher than those obtained with EIE-LVC kit (Se and Sp = 90.01%; 95% CI 82.54 to 95.15%). Bi-epitope ELISA was also able to discriminate *Leishmania*-infected animals from animals infected with *T. cruzi*, however EIE-LVC kit show cross-reactivity with *T. cruzi*-infected sera (Fig 3A and 3B).

Discussion

LdVFA2 protein was identified as an important candidate for vaccine development against visceral leishmaniasis [11, 13, 25].

B-cell epitope in LdVFA2 was previously located within the repetitive units using sera from recLdVFA2-vaccinated mice and with the LdVFA2-specific monoclonal antibody [10]. A recent study has demonstrated that humoral response to recLdVFA2 in dogs is associated, presumably, with protective immunity against *Leishmania* spp. parasites [26], however, B-cell epitope analysis with antibodies elicited in dogs were not conducted. Thus, researches for diagnosis or vaccines applied to dogs could be eventually improved by the identification of B-cell epitope mapping of LdVFA2, using homologous antibodies. In this way, the present study,

Table 2. Seventy-five overlapping decapentapeptides synthesized corresponding to the complete amino acid sequence of A2 *Leishmania sp* protein.

Peptide number	Peptide Sequence	Color intensity (mAb) (pAbs)		Peptide number	Peptide sequence	Color intensity (mAb) (pAbs)	
1	MKIRSVRPLVLLVC	11.10	96.13	39	VGPQSVGPLSVGPQA	49.95	20.22
2	RSVRPLVLLVCVAA	13.43	84.25	40	QSVGPLSVGPQAVGP	39.85	22.07
3	RPLVLLVCVAAVLA	5.34	99.79	41	GPLSVGPQAVGPLSV	63.11	20.82
4	VLLVCVAAVLALSA	6.95	87.53	42	SVGPQAVGPLSVGPQ	52.71	16.86
5	LVCVAAVLALSASAE	10.62	80.68	43	PQAVGPLSVGPQSVG	29.71	19.54
6	VAAVLALSASAEPHK	12.36	36.52	44	VGPLSVGPQSVGPLS	49.13	25.01
7	VLALSASAEPHKAAV	7.94	26.89	45	LSVGPQSVGPLSVGP	63.51	26.16
8	LSASAEPHKAAVDVG	10.69	19.40	46	GPQSVGPLSVGPQAV	26.72	17.06
9	SAEPHKAADVGLS	22.77	31.91	47	SVGPLSVGPQAVGFL	60.71	29.32
10	PHKAAVDVGLSVGP	70.96	32.71	48	FLSVGPQAVGFLSVG	100.00	35.31
11	AAVDVGLSVGPQSV	26.67	20.70	49	VGPQAVGPLSVGPQS	71.15	36.22
12	DVGLSVGPQSVGFL	49.65	29.34	50	QAVGPLSVGPQSVGP	55.11	26.75
13	PLSVGPQSVGFLSVG	85.37	29.66	51	GPLSVGPQSVGFLSV	65.83	22.91
14	VGPQSVGFLSVGPQA	46.38	24.66	52	SVGPQSVGFLSVGPQ	98.62	30.74
15	QSVGPLSVGPQAVGP	43.93	22.79	53	PQSVGFLSVGPQSVG	69.84	29.47
16	GPLSVGPQAVGPLSV	44.48	25.30	54	VGFLSVGPQSVGFLS	79.79	24.82
17	SVGPQAVGFLSVGPQ	61.76	24.81	55	LSVGPQSVGFLSVGS	42.65	16.10
18	PQAVGFLSVGPQSVG	54.83	24.64	56	GPQSVGFLSVGSQSV	32.33	21.36
19	VGFLSVGPQSVGFLS	56.06	25.01	57	SVGFLSVGSQSVGFL	37.82	20.93
20	LSVGPQSVGFLSVGP	67.72	25.85	58	FLSVGSQSVGFLSVG	56.15	26.50
21	GPQSVGFLSVGPQAV	28.75	21.04	59	VGSQSVGFLSVGPQS	38.19	24.80
22	SVGFLSVGPQAVGFL	57.10	32.53	60	QSVGFLSVGPQSVGP	35.07	18.11
23	FLSVGPQAVGFLSVG	98.23	34.70	61	GPLSVGPQSVGFLSV	47.65	19.07
24	VGPQAVGFLSVGPQS	51.55	25.26	62	SVGPQSVGFLSVGPQ	58.78	21.28
25	QAVGFLSVGPQSVGP	55.71	27.91	63	PQSVGFLSVGPQSVG	38.64	21.87
26	GPLSVGPQSVGFLSV	65.37	18.03	64	VGFLSVGPQSVGFLS	44.45	23.09
27	SVGPQSVGFLSVGFL	89.87	27.24	65	LSVGPQSVGFLSVGP	54.45	23.85
28	PQSVGFLSVGFLSVG	79.92	24.14	66	GPQSVGFLSVGPQSV	30.19	18.21
29	VGFLSVGFLSVGPQS	74.04	26.81	67	SVGFLSVGPQSVGFL	50.46	25.96
30	LSVGLSVGPQSVGP	42.24	25.79	68	FLSVGPQSVGFLSVG	74.70	31.42
31	GPLSVGPQSVGFLSV	67.67	22.16	69	VGPQSVGFLSVGPQS	44.99	20.80
32	SVGPQSVGFLSVGSQ	42.65	18.23	70	QSVGFLSVGPQSVGP	31.64	24.81
33	PQSVGFLSVGSQSVG	26.41	20.62	71	GPLSVGPQSVGFLSV	49.29	25.41
34	VGFLSVGSQSVGFLS	35.39	19.50	72	SVGPQSVGFLSVGPQ	51.17	18.40
35	LSVGSQSVGFLSVGP	58.37	25.74	73	PQSVGFLSVGPQSV	58.78	32.72
36	GSQSVGFLSVGPQSV	36.10	21.20	74	VGFLSVGPQSVDVSP	53.32	27.61
37	SVGFLSVGPQSVGFL	34.83	24.31	75	FLSVGPQSVDVSPVS	97.48	32.82
38	FLSVGPQSVGFLSVG	83.90	27.51				

<https://doi.org/10.1371/journal.pntd.0005562.t002>

provided mapping of continuous B-cell epitopes, using sera from *recLdVFA2* immunized dogs and a mouse monoclonal antibody (anti-*recLdVFA2* mAb), by performing a systematic 15 mer peptide-scan along the complete *recLdVFA2* sequence.

Anti-*recLdVFA2* dog sera recognize peptides derived from the non-repetitive N-term region of *recLdVFA2*. Immunoreactivity of peptides bound to membranes reveals five sequences containing epitopes close to this region (peptides 1–15, 4–18, 7–21, 10–24 and 13–27). B-cell epitopes within the repetitive C-term region were not identified. On the other hand using an anti-*LdVFA2* mAb, B-cell epitopes in *recLdVFA2* were located solely within the repetitive units, which is consistent with a previous report [27]. The two most reactive peptides have two repetitions of the sequence PLSVG (spots 23 and 48: region 67–81 and 122–135). This sequence PLSVGPQAVGP LSVG is repeated four times in the complete sequence of the *LdVFA2* protein. Spots 13 and 38




Fig 1. Binding pattern of a dog polyclonal antibodies with the overlapping peptides. (A) Reactivity of 15-mer overlapping peptides derived from the amino acid sequence of A2 protein with dog serum (pAb). Peptides were prepared by the Spot method on cellulose membranes (Section 2) and dogs anti-A2 antibody binding (serum diluted 1:100) to cellulose-bound peptides was detected by an alkaline phosphatase-coupled anti-dog antibody (diluted 1:2000). (B) Percentage of reactivity of each peptide recognized by dog antisera A2 protein. The reactive peptides were: MKIRSVRPLVLLVC, RSVRPLVLLVCVAA, RPLVLLVCVAAVLA, VLLVCVAAVLALSA and LVCVAAVLALSASAE.

<https://doi.org/10.1371/journal.pntd.0005562.g001>

are also very reactive and have only one substitution in its sequence (alanine for serine), when compared to the spots 23 e 48. Thus, residues PLSVG-(X)₅-PLSVG are key contributors to the antigenic recognition of the peptide by specific mAb. However, spots 27 and 75 have only one sequence PLSVG and are also very reactive. We do not have a clear explanation for this observation.

A *recLdVFA2* derived chimeric peptide based on mapped continuous B-cell epitopes MKIRSVRPLVLLVC and PLSVGPQAVGPLSVG was synthesized and used in ELISA




Fig 2. Binding pattern of a murine monoclonal antibodies with the overlapping peptides. (A) Reactivity of 15-mer overlapping peptides derived from the amino acid sequence of A2 protein with a murine monoclonal antibody (mAb). The mAb binding (concentration in 2.0 µg/mL) to cellulose-bound peptides was detected by an alkaline phosphatase-coupled anti-dog antibody (diluted 1:2000). (B) Percentage of reactivity of each peptide recognized by anti-A2 protein murine mAb. The more reactive peptide was PLSVGPQAVGPLSVG.

<https://doi.org/10.1371/journal.pntd.0005562.g002>

experiments for serum diagnosis of CVL. In fact, we prepared 34-mer *recLdVFA2*-derived peptides containing a sequence of 15 residues from one part of *recLdVFA2* linked by a Gly-Gly dipeptide to a pentadeca sequence from another part of the protein, in order to bring together regions that are apart from each other in the linear sequence of the recombinant peptide. Glycine spacers' separating the peptides increases their recognition by antibodies by providing a better exposition of chains for interaction [28]. N-ter and C-ter extremities were capped respectively by a Lysine and a Cysteine residue to allow further specific coupling to protein carriers (not used in this work). The 34-mer bi-epitope peptide was used as coating antigen in an ELISA format.

Accurate diagnosis of canine leishmaniasis is essential towards a more efficient control of this zoonosis, but is no yet achieved due to the high incidence of serological cross-reactions, mainly with other tripanosomatids antigens in canine serum samples [29]. The use of synthetic peptides [30, 31] as antigens in diagnosis of canine leishmaniasis may limit cross reactivity. It would also circumvent reliance on parasite extracts, which are not easy to reproducibly produce, and thus may help assay standardization. In the present work, encouraging results were obtained using the synthetic bi-epitope as coating antigen. Bi-epitope ELISA diagnostic test showed better sensitivity and specificity than the EIE-LVC kit, which is considered the gold standard for CVL diagnosis and high degree of agreement with parasitological techniques for the leishmaniasis diagnosis. These results using peptides selected by a peptide scanning




Fig 3. Comparison of ELISA reactivity and ROC curves obtained from canine sera against recLdVFA2 and antigen of EIE-LVC kit. (A) ELISA reactivity obtained from dogs sera against EIE-LVC kit. (B) ELISA reactivity obtained from dogs sera against recLdVFA2 bi-epitope. The ELISA was performed with groups of 101 uninfected dogs (NI), 101 infected with *Leishmania* (I) and 10 infected with *Trypanosoma cruzi* (TC). (C) ROC curve obtained from EIE-LVC kit. (D) ROC curve obtained from recLdVFA2 bi-epitope. The ROC curves were used to determine cut-off, sensitivity, specificity and AUC. Data variation expressed as standard error. *Cut-off obtained by ROC curve and #Cut-off obtained according to kit manufacturer.

<https://doi.org/10.1371/journal.pntd.0005562.g003>

method showed a better performance than others studies with synthetic peptides for CVL immunodiagnosis tests [32, 33, 30, 34].

In conclusion, data presented in the current study suggest that it is feasible to map B-cell epitopes from an overlapping peptide library covering the full length of recLdVFA2 and to further use the selected peptides in combination to diagnose canine visceral leishmaniasis. Further studies using sera of dogs from endemic areas (with high and low CVL prevalence) are obviously required to determine the use of these antigens for field control of CVL. Bi-epitope is derived from A2 protein and Leish-Tec vaccine is produced from this protein, so if it used in a diagnostic test of CVL, it may not be able to discriminate infected animals from those vaccinated. This antigen can be used in regions where this vaccine is not used and can also be combined with other antigenic epitopes that can minimize this fact. Finally, our work suggests that synthetic peptide-based ELISA strategy may be useful for the development of a sensitive and highly specific serodiagnosis for CVL or other parasitic diseases.

Table 3. Diagnostic performance for recLdVFA2 and EIE-LVC kit.

	Test	
	recLdVFA2	EIE-LVC kit
Cut-off	0.4710*	0.0665*
AUC	0.9987	0.9601
ACC	0.9851	0.9010
Se (95% CI)	98.02% (93.03–99.76%)	0.9010% (82.54–95.15%)
Sp (95% CI)	99.01% (94.61–99.97%)	0.9010% (82.54–95.15%)
PPV	0.99	0.90
NPV	0.98	0.90
Kappa (95% CI)	0.97(0.937–1.00) (very good)	0.802 (0.720–0.884) (very good) [#]

*Cut-off obtained by ROC curve

[#]With cut-off obtained according to kit manufacturer the agreement is “good”

Agreement was calculated using parasitological assays as gold standard test.

Abbreviations: AUC: area under curve; ACC: accuracy; Se: sensitivity; Sp: specificity; PPV: positive predictive value; NPV: negative predictive value; CI: confidence interval.

<https://doi.org/10.1371/journal.pntd.0005562.t003>

Supporting information

S1 STARD Checklist.

(DOCX)

Acknowledgments

We thank Dr. Eduardo Antônio Ferraz Coelho, Colégio Técnico da Universidade Federal de Minas Gerais, Brasil, for providing anti-recLdVFA2 protein dogs sera, Dra. Andreza Pain Marcelino, Fundação Ezequiel Dias–FUNED, Minas Gerais, Brasil, for providing an EIE-LVC kit Bio-Manguinhos and Dr. Greg Matlashewski, McGill University, Quebec, Canada, for kindly providing us with anti- *L. donovani* virulence factor A2 murine monoclonal antibody (mAb-anti recLdVFA2).

Author Contributions

Conceived and designed the experiments: TMM EHR OBR DCB RTF CCO.

Performed the experiments: TMM EHR FCO LDCL CMTM.

Analyzed the data: TMM EHR RTF CCO.

Contributed reagents/materials/analysis tools: DCB RTF.

Wrote the paper: TMM EHR CCO.

References

1. World Health Organization. Control of Leishmaniasis. Technical Reports Series 1990, v. 793, p. 50–52.
2. Pace D. Leishmaniasis. *J. Infect.* 2014; 69(1):S10–8
3. Cabral M, O’Grady JE, Gomes S, Sousa JC, Thompson H, Alexander J. The immunology of canine leishmaniasis: strong evidence for a developing disease spectrum from asymptomatic dogs. *Vet. Parasitol.* 1998; 76(3): 173–180. PMID: [9615951](https://pubmed.ncbi.nlm.nih.gov/9615951/)
4. Sideris V, Papadopoulou G, Dotsika E, Karagouni E. Asymptomatic canine leishmaniasis in Greater Athens area, Greece. *Eur. J. Epidemiol.* 1999, 15(3): 271–276. PMID: [10395058](https://pubmed.ncbi.nlm.nih.gov/10395058/)

5. Regina-Silva S, Fortes-Dias CL, Michalsky EM, França-Silva JC, Quaresma PF, Lima ACVMR, et al. Evaluation of parasitological examination, kDNA polymerase chain reaction and rK39-based immunochromatography for the diagnosis of visceral leishmaniasis in seropositive dogs from the screening-culling program in Brazil. *Rev. Soc. Bras. Med. Tropical* 2014; 47(4):462–468.
6. Charest H and Matlashewski G. Developmental gene expression in *Leishmania donovani*: differential cloning and analysis of an amastigote-stage-specific gene. *Mol Cell Biol* 1994; 14: 2975–2984. PMID: [7545921](#)
7. Zhang WW and Matlashewski G. Loss of virulence in *Leishmania donovani* deficient in an amastigote-specific protein, A2. *Proc Natl Acad Sci USA* 1997; 94: 8807–8811. PMID: [9238059](#)
8. Mac Call LI and Matlashewski G. Localization and induction of the A2 virulence factor in *Leishmania*: evidence that A2 is a stress response protein. *Molecular Microbiology* 2010; 77(2): 518–530. <https://doi.org/10.1111/j.1365-2958.2010.07229.x> PMID: [20497497](#)
9. Zhang WW, Charest H, Ghedin E, Matlashewski G. Identification and overexpression of the A2 amastigote-specific protein in *Leishmania donovani*. *Mol Biochem Parasitol* 1996; 78: 79–90. PMID: [8813679](#)
10. Ghedin E, Zhang WW, Charest H, Sundar S, Kenney RT, Matlashewski G. Antibody response against a *Leishmania donovani* amastigote-stage-specific protein in patients with visceral leishmaniasis. *Clin Diagn Lab Immunol* 1997; 4: 530–535. PMID: [9302200](#)
11. Ghosh A, Labrecque S, Matlashewski G. Protection against *Leishmania donovani* infection by DNA vaccination: increased DNA vaccination efficiency through inhibiting the cellular p53 response. *Vaccine* 2001; 19: 3169–3178. PMID: [11312013](#)
12. Zanin FH, Coelho EA, Tavares CA, Marques-da-Silva EA, Silva Costa MM, Rezende SA, et al. Evaluation of immune responses and protection induced by A2 and (NH) DNA vaccines against *Leishmania chagasi* and *Leishmania amazonensis* experimental infections. *Microbes Infect.* 2007; 9: 1070–1077. <https://doi.org/10.1016/j.micinf.2007.05.012> PMID: [17644455](#)
13. Fernandes AP, Costa MM, Coelho EA, Michalick MS, de Freitas E, Melo MN, et al. Protective immunity against challenge with *Leishmania (Leishmania) chagasi* in beagle dogs vaccinated with recombinant A2 protein. *Vaccine* 2008; 26: 5888–5895. <https://doi.org/10.1016/j.vaccine.2008.05.095> PMID: [18786587](#)
14. Grimaldi G Jr, Teva A, Porrozzi R, Pinto MA, Marchevsky RS, Rocha MG, et al. Clinical and parasitological protection in a *Leishmania infantum*-macaque model vaccinated with adenovirus and the recombinant A2 antigen. *PLoS Negl. Trop. Dis.* 2014; 8(6): e2853. <https://doi.org/10.1371/journal.pntd.0002853> PMID: [24945284](#)
15. Carvalho FA, Charest H, Tavares CA, Matlashewski G, Valente EP, Rabello A, et al. Diagnosis of American visceral leishmaniasis in humans and dogs using the recombinant *Leishmania donovani* A2 antigen. *Diagn. Microbiol. Infect. Dis.* 2002; 42: 289–295.
16. Frank R. The SPOT-synthesis technique. Synthetic peptide arrays on membrane supports-principles and applications. *J Immunol Methods* 2002; 267: 13–26. PMID: [12135797](#)
17. Chavez-Olortegui C, Molina F, Granier C. Molecular basis for the cross-reactivity of antibodies elicited by a natural anatoxin with alpha- and beta-toxins from the venom of *Tityus serrulatus* scorpion. *Mol. Immunol.* 2002; 38(11): 867–876. PMID: [11922945](#)
18. Laune D, Molina F, Ferrieres G, Villard S, Bes C, Rieunier F, et al. Application of the Spot method to the identification of peptides and amino acids from the antibody paratope that contribute to antigen binding. *J Immunol Methods* 2002; 267: 53–70. PMID: [12135800](#)
19. Frank R. Spot-Synthesis: an easy technique for the positionally addressable, parallel chemical synthesis on a membrane support. *Tetrahedron* 1992; 48: 9217–9232.
20. Merrifield RB. Solid-phase peptide synthesis. *Adv. Enzimol. Relat. Areas Mol. Biol.* 1969; 32: 221–296.
21. Machado de Avila RA, Stransky S, Velloso M, Castanheira P, Schneider FS, Kalapothakis E, et al. Mimotopes of mutalysin-II from *Lachesis muta*, snake venom induce hemorrhage inhibitory antibodies upon vaccination of rabbits. *Peptide* 2011; 32: 1640–1646.
22. Chavez-Olortegui C, Amaral DA, Rochat H, Diniz C, Granier C. *In vivo* protection against scorpion toxins by liposomal immunization. *Vaccine* 1991; 9: 907–910. PMID: [1811376](#)
23. Linnet K, Bossuyt PM, Moons KG, Reitsma JB. Quantifying the accuracy of a diagnostic test or marker. *Clin. Chem.* 2012; 58: 1292–1301. <https://doi.org/10.1373/clinchem.2012.182543> PMID: [22829313](#)
24. Molina F, Laune D, Gougat C, Pau B, Granier C. Improved performances of spot multiple peptide synthesis. *Pept Res* 1996; 9: 151–155. PMID: [8875595](#)
25. Coelho EA, Tavares CA, Carvalho FA, Chaves KF, Teixeira KN, Rodrigues RC, et al. Immune responses induced by the *Leishmania (Leishmania) donovani* A2 antigen, but not by the LACK antigen, are protective against experimental *Leishmania (Leishmania) amazonensis* infection. *Infect Immun* 2003; 71:3988–3994. <https://doi.org/10.1128/IAI.71.7.3988-3994.2003> PMID: [12819086](#)

26. Testasicca MCS, Santos MS, Machado LM, Serufo AV, Doro D, Avelar D, et al. Antibody responses induced by Leish-Tec, an A2-based vaccine for visceral leishmaniasis, in a heterogeneous canine population. *Veterinary Parasitology* 2014; 204: 169–176. <https://doi.org/10.1016/j.vetpar.2014.04.025> PMID: [24863572](https://pubmed.ncbi.nlm.nih.gov/24863572/)
27. Resende DM, Caetano BC, Dutra MS, Penido MLO, Abrantes CF, Verly RM, et al. Epitope mapping and protective immunity elicited by adenovirus expressing the *Leishmania* amastigote specific A2 antigen: Correlation with IFN- γ and cytolytic activity by CD8⁺ T cells. *Vaccine* 2008; 26: 4585–4593. <https://doi.org/10.1016/j.vaccine.2008.05.091> PMID: [18588933](https://pubmed.ncbi.nlm.nih.gov/18588933/)
28. Bialek M, Grabowski S, Kaminski Z, Kaca W. Synthetic peptides mimicking antigenic epitope of *Helicobacter pylori* urease. *Acta Biochim. Pol.* 2006; 53(1): 83–86. PMID: [16496040](https://pubmed.ncbi.nlm.nih.gov/16496040/)
29. Troncarelli MZ, Camargo JB, Machado JG, Lucheis SB, Langoni H. *Leishmania* spp and/or *Trypanosoma cruzi* diagnosis in dogs from endemic and nonendemic areas for canine visceral leishmaniasis. *Vet. Parasitol.* 2009; 164(2–4): 118–123. <https://doi.org/10.1016/j.vetpar.2009.06.027> PMID: [19625128](https://pubmed.ncbi.nlm.nih.gov/19625128/)
30. Costa MM, Penido M, Santos MS, Doro D, Freitas E, Michalick MSM, et al. Improved canine and human visceral leishmaniasis immunodiagnosis using combinations of synthetic peptides in enzyme-linked immunosorbent assay. *PLoS Negl. Trop. Dis.* 2012; 6(5): e1622. <https://doi.org/10.1371/journal.pntd.0001622> PMID: [22629475](https://pubmed.ncbi.nlm.nih.gov/22629475/)
31. Toledo-Machado CM, de Avila RA, NGuyen C, Granier C, Bueno LL, Carneiro CM, et al. Immunodiagnosis of canine visceral leishmaniasis using mimotope peptides selected from phage displayed combinatorial libraries. *Biomed. Res. Int.* 2015; 401509.
32. Faria AR, Costa MM, Giusta MS, Grimaldi G Jr, Penido MLO, Gazzinelli RT, Andrade HM. High-throughput analysis of synthetic peptides for the immunodiagnosis of canine visceral leishmaniasis. *PLoS Negl. Trop. Dis.* 2011, 5(9): e1310a.
33. Ramos-Jesus J, Carvalho KA, Fonseca RAS, Oliveira GGS, Melo SMB, Alcântara-Neves NM, Dutra RF. A piezoelectric immunosensor for *Leishmania chagasi* antibodies in canine serum. *Anal. Bioanal. Chem.* 2011, 401:917–925 <https://doi.org/10.1007/s00216-011-5136-7> PMID: [21667350](https://pubmed.ncbi.nlm.nih.gov/21667350/)
34. Menezes-Souza D, Mendes TAOM, Gomes MS, Reis-Cunha JL, Nagem RAP, Carneiro CM et al. Epitope mapping of the HSP83.1 protein of *Leishmania braziliensis* discloses novel targets for immunodiagnosis of tegumentary and visceral clinical forms of Leishmaniasis. *Clin. Vaccine Immunol.* 2014, 21(7): 917–925.